

The management system of

Alesi Surgical Limited

Cardiff Medicentre, Heath Park, Cardiff, CF14 4UJ, UK

has been assessed and certified as meeting the requirements of

Directive 93/42/EEC on medical devices, Annex II (excluding Section 4)

For the following products

Ultravision Electrostatic Precipitator and Sterile Trocars.

Where the above scope includes class III medical device(s), a valid EC Design Examination Certificate according to Annex II (Section 4) is a mandatory requirement for each device in addition to this certificate to place that device on the market.

This certificate is valid from 18 July 2016 until 03 January 2019
and remains valid subject to satisfactory surveillance audits.

Re certification audit due before 01 October 2016

Issue 4. Certified since 03 January 2014

Certification is based on reports numbered GB/PC 230872

Authorised by

SGS United Kingdom Ltd, Notified Body 0120

202B Worle Parkway, Weston-super-Mare, BS22 6WA UK
t +44 (0)1934 522917 f +44 (0)1934 522137 www.sgs.com

SGS CE 02 0215

Page 1 of 1

